

INTELLIGENT Gas

Marcelo T. de Alvear 4304/06 – Ciudadela (B 1702CFZ)
Tel. : (011) 4647-1950 / 2056-8100 Líneas Rotativas - Skype: ibsanitarios
www.intelligentgas.com.ar / www.ibsanitarios.com.ar / www.calefactorescalden.com.ar

CERTIFICATE OF ANALYSIS

Date: February 22, 2023
Order Number: Gas-3102
Lot Number: 304-402657444-1

Customer: DEL SUR INDUSTRIAL SRL
Use Before: 02/22/2025

<u>Component</u>	<u>Requested Concentration</u>	<u>Analytical Result (+/- 2%)</u>
Carbon Monoxide	100 PPM	101 PPM
Methane	2.5% vol.	2.51% vol.
Hydrogen Sulfide	25 PPM	26 PPM
Oxygen	15% vol.	15.30% vol.
Nitrogen	Balance	Balance

Cylinder Size: 2.0 Cu. Ft.
Contents: 58 Liter

Valve: 5/8" -18UNF
Pressure: 500 psig

Product composition verified by direct comparison to calibration standards traceable to N.I.S.T. weights and/or N.I.S.T. Gas Mixture reference materials.

Analyst:


Glenn Velez